



Certificate of Compliance

Certificate: 2595034 (LR 53988_C)

Master Contract: 151079

Project: 2595034

Date Issued: April 4, 2013

Issued to: Endress + Hauser GmbH + Co. KG
Hauptstrasse 1
Maulburg, Baden-Wuerttemberg 79690
Germany
Attention: Dr. Arno Götz

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Lino Menezes

Issued by: Lino Menezes

PRODUCTS

CLASS 2252 05 - PROCESS CONTROL EQUIPMENT

CLASS 2252 85 - PROCESS CONTROL EQUIPMENT - Certified to US Standards

Prosonic Model T FMU30- (followed by additional numbers and / or letters denoting specific features) is a permanently connected, non-contact Level Measuring Device. Depending on the sensor (1-1/2" or 2" Sensor), the measuring range is up to 8 m (26 ft) in fluids and up to 3,5 m (11 ft) in bulk solids. Ultrasonic level measurement, stationary mounted, CSA Enclosure Type 6P / IP68. An Electronic module FEU31 is used with ratings as follows: 14...35 VDC, 0.8W; 4...20 mA, two wire output; maximum ambient temperature 60 °C.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Second Edition)

ANSI/UL Std. No. 61010-1, 2nd Edition - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Second Edition)



Supplement to Certificate of Compliance

Certificate: 2595034

Master Contract: 151079

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
2595034	Apr 4, 2013	Original Certification.